



Bid Number/बोली क्रमांक (बिड संख्या):
GEM/2024/B/5283864
Dated/दिनांक : 19-09-2024

Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	04-10-2024 11:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	04-10-2024 11:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	180 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Ministry Of Power
Department Name/विभाग का नाम	Power Grid Corporation Of India Limited
Organisation Name/संगठन का नाम	Power Grid Corporation Of India Limited
Office Name/कार्यालय का नाम	Corporate Office Gurgaon
Total Quantity/कुल मात्रा	5
Item Category/मद केटेगरी	Insulation Tester (Q3) , Digital Multimeter as per IS 9000 (Q3)
MSE Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से एमएसई छूट	No
Startup Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से स्टार्टअप छूट	No
Document required from seller/विक्रेता से मांगे गए दस्तावेज़	Certificate (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Do you want to show documents uploaded by bidders to all bidders participated in bid?/	Yes
Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया	No
Type of Bid/बिड का प्रकार	Two Packet Bid
Primary product category	Insulation Tester
Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय	3 Days
Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No

Bid Details/बिड विवरण

Estimated Bid Value/अनुमानित बिड मूल्य	342214
Evaluation Method/मूल्यांकन पद्धति	Total value wise evaluation

EMD Detail/ईएमडी विवरण

Required/आवश्यकता	No
-------------------	----

ePBG Detail/ईपीबीजी विवरण

Required/आवश्यकता	No
-------------------	----

MII Purchase Preference/एमआईआई खरीद वरीयता

MII Purchase Preference/एमआईआई खरीद वरीयता	No
--	----

MSE Purchase Preference/एमएसई खरीद वरीयता

MSE Purchase Preference/एमएसई खरीद वरीयता	Yes
---	-----

1. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

2. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

Insulation Tester (4 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
PERFORMANCE	Type	Hand Held
	Measurement Range	0.01 M Ohm to 10 G Ohm
	Measurement Accuracy	2.0 (percent)
	Measurement Resolution	0.01 (megaOhm)
	Test Voltage	1000.0 (Volt)
	Test Current	1.0 (milliAmpère)
OTHER MEASUREMENT	AC/DC Voltage Measurement (Volt)	600
	AC/DC Voltage Measurement Resolution (Volts)	0.1
	AC/DC Voltage Measurement Accuracy (%)	2.0
FEATURES	Live Circuit Detection	inhibit test terminal voltage >30 V prior to initialization of test
SAFETY	Safety rating	Category III 1000 V of IEC 61010
POWER REQUIREMENT	Power Requirements	Battery Operated
	Type of Battery	Non- Rechargeable (Alkaline)
WARRANTY	Warranty	3.0 (year)
OPERATING CONDITION	Minimum Operating Temperature	-10.0, -20.0 (degree Celsius)
	Maximum Operating Temperature	60.0 (degree Celsius)
	Operating Humidity (RH)	90.0 (percent)

Additional Specification Parameters - Insulation Tester (4 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Auto discharge time	Auto discharge time shall be less than 0.5 sec

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्र

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्र	Delivery Days/डिलीवरी के दिन
1	Brij Bhushan Singh	211008,Powergrid Corporation of India Limited, Allahabad Substation, Village-Sarangapur, Post Office-Dhandupur, Rewa Road, Allahabad (Prayagraj), Uttar Pradesh-211008	1	90
2	Raju Singh	273209,POWERGRID CORPORATION OF INDIA LIMITED, Gorakhpur Substation, Gorakhpur 400/220KV S/S, Sector-23, Gida Sahajanwa, Gorakhpur, Uttar Pradesh-273209	1	90
3	Ahamed Hasan	209311,Power Grid Corporation of India Limited, Northern Region Transmission System-III, 765/400 kV GIS Substation , RANIA, Kanpur Dehat, Pin - 209311 (UP)	1	90
4	Sandeep Kothari	262543,Power Grid Corporation of India Limited Northern Region-III, 400/220 kV (under construction) GIS Substation, V. Bagri hatt, P. Askot Dist.- Pithoragarh, Uttarakhand, Pin-262543	1	90

Digital Multimeter As Per IS 9000 (1 pieces)**Technical Specifications/तकनीकी विशिष्टियाँ**

[* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
MEASUREMENTS	DC Voltage Range	1000.0 (Volt)
	DC Voltage Accuracy	0.1 (percent)
	DC Voltage resolution	0.1 (milliVolt)
	AC Voltage Range	1000.0 (Volt)
	AC Voltage Accuracy	0.1 (percent)
	AC Voltage resolution	0.1 (milliVolt)
	DC Current Range	10.0 (Ampère)
	DC Current Accuracy	1.0 (percent)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	DC Current resolution	0.01 (milliAmpère)
	AC Current Range	10.0 (Ampère)
	AC Current Accuracy	1.5 (percent)
	AC Current resolution	0.1 (milliAmpère)
	Resistance Range	50.0 (megaOhm)
	Resistance Accuracy	1.0 (percent)
	Resistance resolution	0.1 (Ohm)
	Capacitance Range (Micro farad)	.01 to 1000
	Capacitance Accuracy (%)	1.5
	Capacitance resolution (nF)	1
	Frequency Range (KHz)	100
	Frequency Accuracy (%)	0.1
	Frequency resolution (Hz)	0.01
	Temperature Range (Degree C)	-40 to 400
	Temperature Accuracy (%)	1.5
	Temperature resolution (Degree C)	1
FEATURES	Type	Hand Held
	Display (Digits)	3.5
	Measurement Type	True RMS
	Counts	4000, 6000, 50000, 10000, 20000, 30000, 40000, 60000, 70000, 80000, 90000, 100000, 200000, 2000000, 2000 Or higher
OTHER FEATURES	Continuity test and Diode test	Available
SAFETY	Safety Rating	Category IV 600 V of IEC 61010
	Safety Certification	UL, CE, CSA, TUV, VDE
POWER REQUIREMENTS	if Battery Operated, then Type of Battery	Non- Rechargeable (Alkaline)
WARRANTY	Warranty	5.0 (year)
OPERATING CONDITION	Minimum Operating Temperature	-10.0 (degree Celsius)
	Maximum Operating Condition	50.0 (degree Celsius)
	Operating Humidity (RH)	95.0, 90.0 (percent)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
TEST REPORT DETAILS	The equipment supplied with calibration certificate	YES, NA

Additional Specification Parameters - Digital Multimeter As Per IS 9000 (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Accessories	test lead should be supplied by bidder

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Gaurav Gupta	262501,Power Grid Corporation of India Limited Northern Region-III, 220/132 KV Pithoragarh Substation,Vill.- Mostamanu,P.O- Chandak,Distt.- Pithoragarh,Uttarakhand.- 262501	1	90

Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें

1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

2. Generic

- The Seller shall not assign the Contract in whole or part without obtaining the prior written consent of buyer.
- The Seller shall not sub-contract the Contract in whole or part to any entity without obtaining the prior written consent of buyer.
- The Seller shall, notwithstanding the consent and assignment/sub-contract, remain jointly and severally liable and responsible to buyer together with the assignee/ sub-contractor, for and in respect of the due

performance of the Contract and the Sellers obligations there under.

3. **Buyer Added Bid Specific ATC**

Buyer Added text based ATC clauses

Buyer has uploaded Additional Terms & Conditions (ATC) Documents under Additional SLA head/Buyer added bid specific ATC, all Bidders are to upload the same document under Technical Experience head on GeM Portal during submission of Bid, completely filled with required ATC document and technical specification and other necessary information. If Bidder fails to submit the ATC document and technical specification, their bid shall be considered non-responsive.

Terms & Conditions under the uploaded ATC Documents are the superseding Conditions on all other Terms & Conditions referred in the Bid Document

4. **Certificates**

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

5. **Buyer Added Bid Specific ATC**

Buyer uploaded ATC document [Click here to view the file.](#)

Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the

case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।

---Thank You/धन्यवाद---